



Iron, Ferrous, Test Kit IR-18C (2667200)


DOC326.97.00063

Test preparation


CAUTION: ⚠ *Review the Safety Data Sheets (MSDS/SDS) for the chemicals that are used. Use the recommended personal protective equipment.*

- Analyze samples immediately after collection.
- Put the color disc on the center pin in the color comparator box (numbers to the front).
- Use the indoor light color disc when the light source is fluorescent light. Use the outdoor light color disc when the light source is sunlight.
- Rinse the tubes with sample before the test. Rinse the tubes with deionized water after the test.
- If the color match is between two segments, use the value that is in the middle of the two segments.
- If the color disc becomes wet internally, pull apart the flat plastic sides to open the color disc. Remove the thin inner disc. Dry all parts with a soft cloth. Assemble when fully dry.
- Undissolved reagent does not have an effect on test accuracy.
- To verify the test accuracy, use a standard solution as the sample.
- This test kit measures ferrous iron. To determine ferric iron (Fe^{3+}), subtract the ferrous iron result from a total iron test.


Test procedure—Iron, ferrous (0–7 mg/L Fe^{2+})




1. Fill a tube to the first line (5 mL) with sample.




2. Put the tube into the left opening of the color comparator box.




3. Fill the vial to the 25-mL mark with sample.




4. Add one Ferrous Iron Reagent Powder Pillow to the vial.




5. Swirl to mix. A orange color develops if ferrous iron is in the sample.




6. Wait 3 minutes.




7. Fill a second tube to the first line (5 mL) with the prepared sample.



8. Put the second tube into the color comparator box.



9. Hold the color comparator box in front of a light source. Turn the color disc to find the color match.



10. Read the result in mg/L in the scale window.

Replacement items

Description	Unit	Item no.
Ferrous Iron Reagent Powder Pillows	100/pkg	103769
Color disc, iron, indoor light, 0–7 mg/L	each	9261000
Color disc, iron, outdoor light, 0–7 mg/L	each	9263700
Color comparator box	each	173200
Plastic viewing tubes, 18 mm, with caps	4/pkg	4660004
Vial with 2, 5, 10, 15, 20 and 25-mL marks	each	219300

Optional items

Description	Unit	Item no.
Caps for plastic viewing tubes (4660004)	4/pkg	4660014
Glass viewing tubes, glass, 18 mm	6/pkg	173006
Stoppers for 18-mm glass tubes and AccuVac Ampuls	6/pkg	173106
Water, deionized	500 mL	27249

